

Date/Time Manual at 15:21:50 August 22, 2010
Range Geo : 10.00 in/s
Record Time 25.0 sec at 1024 sps
Job Number: 5177
Notes

Serial Number BE11100 V 10.10-8.17 MiniMate Plus
Battery Level 6.4 Volts
Unit Calibration October 21, 2009 by Instantel Inc.
File Name M100DDPX.CEO
Post Event Notes



Extended Notes

Measure at concrete culvert beneath road 31' from vibes at station 179.5

	Tran	Vert	Long	
PPV	0.125	0.365	0.310	in/s
ZC Freq	23	27	32	Hz
Time (Rel. to Trig)	0.382	0.006	1.488	sec
Peak Acceleration	0.0795	0.133	0.199	g
Peak Displacement	0.00075	0.00276	0.00191	in
Sensor Check	Disabled	Disabled	Disabled	
Frequency	***	***	***	Hz
Overswing Ratio	***	***	***	

Peak Vector Sum 0.377 in/s at 0.008 sec

USBM RI8507 And OSMRE



Sensor Check